



Strojírenský zkušební ústav, s.p. (Engineering Test Institute, Public Enterprise),
Hudcova 424/56b, 621 00 Brno, Czech Republic

CERTIFICATE

B-32-00336-16

Manufacturer: Arkon Flow Systems, s.r.o.
Berkova 534/92, 612 00 Brno - Královo Pole
Czech Republic
Company ID No.: 27683826

Products: Electromagnetic flowmeters

Type designation: MAGB1
Versions: Compact, Remote

The Engineering Test Institute, Public Enterprise, hereby certifies that the characteristics of the sample of the products concerned have been found conforming to the applicable requirements

for the IP 68 protection rating as per ČSN EN 60529:1993 (at a water level of 1.25 m above the top edge of the sample for the duration of 15 hours).

This Certificate has been issued based on Final Report 32-0261 of 2016-03-23, issued by the Engineering Test Institute, Public Enterprise.

The rules for using the Certificate are specified on Page 2.

Brno, 2016-03-23



Ing. Tomáš Hruška
Director



RULES FOR USING THE CERTIFICATE:

For the purposes of placing the products on the market, this Certificate may only be used provided the normative document specified on Page 1 remains valid.

The Certificate may only be used as a certificate for the products specified on Page 1. This also applies to its use in advertising, promotional and commercial materials. Unauthorised or deceptive use of the Certificate may result in its withdrawal.

It is prohibited to change, amend or overwrite the data contained in the Certificate.

This Certificate may not be used for the products changed without the consent of the Engineering Test Institute, Public Enterprise, in such a way that conformity with the technical requirements specified on Page 1 has been affected.





Strojírenský zkušební ústav, s.p. (Engineering Test Institute, Public Enterprise),
Hudcova 424/56b, 621 00 Brno, Czech Republic

CERTIFICATE

B-32-00337-16

Manufacturer: Arkon Flow Systems, s.r.o.
Berkova 534/92, 612 00 Brno - Královo Pole
Czech Republic
Company ID No.: 27683826

Products: Electromagnetic flowmeters

Type designation: MAGX2

Versions: Compact, Remote

The Engineering Test Institute, Public Enterprise, hereby certifies that the characteristics of the sample of the products concerned have been found conforming to the applicable requirements


for the IP 68 protection rating as per ČSN EN 60529:1993 (at a water level of 1.25 m above the top edge of the sample for the duration of 15 hours).

This Certificate has been issued based on Final Report 32-0261 of 2016-03-23, issued by the Engineering Test Institute, Public Enterprise.

The rules for using the Certificate are specified on Page 2.

Brno, 2016-03-23




Ing. Tomáš Hruška
Director



RULES FOR USING THE CERTIFICATE:

For the purposes of placing the products on the market, this Certificate may only be used provided the normative document specified on Page 1 remains valid.

The Certificate may only be used as a certificate for the products specified on Page 1. This also applies to its use in advertising, promotional and commercial materials. Unauthorised or deceptive use of the Certificate may result in its withdrawal.

It is prohibited to change, amend or overwrite the data contained in the Certificate.

This Certificate may not be used for the products changed without the consent of the Engineering Test Institute, Public Enterprise, in such a way that conformity with the technical requirements specified on Page 1 has been affected.

